

Certification Report

Certified Reference Material

BAM-M603

Zinc powder

September 2019

Coordinator:

Dr. Sebastian Recknagel
Bundesanstalt für Materialforschung und -prüfung (BAM)
Division 1.6 „Inorganic Reference Materials“
Richard-Willstätter-Str. 11
12489 Berlin
Phone: ++49/30/8104 1111
Fax.: ++49/30/8104 1117
E-mail: sebastian.recknagel@bam.de

Summary

This report describes preparation, analysis and certification of zinc powder reference material BAM-M603.

The certified reference material (CRM) is available in the form of powder ($d_{50} = 148 \mu\text{m}$). It is intended for establishing and checking the calibration of wet chemical methods for the analysis of samples of similar materials.

The following mass fractions and uncertainties have been certified:

Element	Mass fraction ¹⁾ in mg/kg	Uncertainty ²⁾ in mg/kg
Pb	15.8	0.5
Ag	1.00	0.09
Cd	1.69	0.12
Cu	3.69	0.21
Fe	2.18	0.14
Ni	0.43	0.05
Tl	3.81	0.23

¹⁾ Unweighted mean value of the means of accepted sets of data (consisting of at least 4 but usually 5 single results), each set being obtained by a different laboratory and/or a different method of measurement.

²⁾ Estimated expanded uncertainty U with a coverage factor of $k = 2$, corresponding to a level of confidence of approx. 95 %, as defined in the Guide to the Expression of Uncertainty in Measurement, (GUM, ISO/IEC Guide 98-3:2008).

The certified values are based on the results of 14 laboratories which participated in the certification inter-laboratory comparison.

The mass fractions of Al, As, Bi, Co, In, Sb, Sn, and V are given for information.

Content

	Page
List of abbreviations	5
1. Introduction.....	6
2. Companies/laboratories involved	6
3. Candidate material	7
4. Homogeneity testing.....	7
5. Stability testing	8
6. Characterisation study	8
6.1 Analytical methods	8
6.2 Analytical results and statistical evaluation.....	10
7. Instructions for users.....	28
8. Metrological Traceability	28
9. Information on and purchase of the CRM.....	29
10. References	29
Annex 1: Calculation of uncertainty contribution of potential inhomogeneity	30

List of abbreviations

(if not explained elsewhere)

CRM	certified reference material
ETAAS	electrothermal atomic absorption spectrometry
ICP-OES	inductively coupled plasma optical emission spectrometry
ICP-MS	inductively coupled plasma mass spectrometry
M	mean value
n	number of accepted data sets
s	standard deviation of an individual data set
S_M	standard deviation of laboratory means
S_{rel}	relative standard deviation
\bar{s}_i	square root of mean of variances of data sets under repeatability conditions
M_i	single result
ICP	ICP-OES (Tables 5 – 19)
U_{ilc}	uncertainty contribution from inter-laboratory comparison
U_{stab}	uncertainty contribution from possible instability
U_{bb}	uncertainty contribution from possible inhomogeneity

1. Introduction

Zinc powder is mainly used in alkaline batteries and in button cells, in thermal spraying for corrosion protection and in the chemical industry. DIN EN 1179 (Primary zinc) defines the maximum contents of the elements Pb, Cd, Fe, Sn, Cu, and Al [1]. For 99.995 % zinc (Z1 acc. to the standard) the following mass fractions are defined:

Pb: max. 30 mg/kg

Cd: max. 30 mg/kg

Fe: max. 20 mg/kg

Sn: max. 10 mg/kg

Cu: max. 10 mg/kg

Al: max. 10 mg/kg

The sum of all six elements must not be higher than 50 mg/kg.

The idea to produce a zinc powder reference material was the outcome of discussions within the German Gesellschaft der Metallurgen und Bergleute e.V. (GDMB), especially of the working group „Zinc“ of the Committee of Chemists within GDMB. The needs are defined by this working group, since the members are potential users of the prepared CRMs. Participating laboratories were recruited from this group. Since all the laboratories participating in this certification project are highly experienced with zinc analysis and had participated in earlier inter-laboratory comparisons, there was no preceding round robin for qualification.

Certification of reference material BAM-M603 was carried out on the basis of ISO 17034 [2] and the relevant ISO-Guides [3, 4].

2. Companies/laboratories involved

Manufacturing of the material

- Grillo AG, Goslar, Germany

Test for homogeneity and stability

- Bundesanstalt für Materialforschung und -prüfung (BAM), Berlin, Germany

Participants in the certification inter-laboratory comparison

- Agilent GmbH AG, Waldbronn, Germany
- AMCO united, Duisburg, Germany
- Boliden Odda, Norway
- Bundesanstalt für Materialforschung und -prüfung (BAM), Berlin, Germany
Division „Inorganic Reference Materials“
Division „Inorganic Trace Analysis“
- Chemad GmbH, Duisburg, Germany
- Inspectorate International Limited, Witham, Essex, United Kingdom
- MKM Mansfelder Kupfer und Messing GmbH, Hettstedt, Germany
- Norzinco GmbH, Goslar, Germany
- revierlabor GmbH, Essen, Germany
- Thyssenkrupp Steel Europe AG, Duisburg, Germany
- TU Clausthal, Institut für Aufbereitung, Deponietechnik und Geomechanik, Clausthal, Germany
- VARTA Consumer Batteries GmbH & Co. KGaA, Dischingen, Germany
- VARTA Microbattery GmbH, Ellwangen, Germany

Statistical evaluation of the data

– Bundesanstalt für Materialforschung und -prüfung (BAM), Berlin, Germany

3. Candidate material

The starting material in form of fine powder was taken from the normal production process within Grillo-Werke Goslar. The particle size was determined by the producer using dry sieve analysis according to ASTM B214 [5].

Tab. 1: Particle size distribution of BAM-M603

range	amount
> 425 µm	0 %
425-250 µm	12 %
250-180 µm	25 %
180-150 µm	11 %
150-106 µm	25 %
106-75 µm	14 %
75-45 µm	9 %
45-25 µm	3 %
< 25 µm	1 %
d 10	66 µm
d 50	148 µm
d 90	284 µm

The powder was homogenised and bottled in BAM using a spinning riffler. In total 495 bottles containing 200 g of powder were obtained from the whole batch.

4. Homogeneity testing

A total of 15 bottled units (see Table 2) of the candidate material were randomly selected for homogeneity testing.

Tab. 2: Bottles analysed for homogeneity testing of BAM-M603

3	114	206	334	417
36	133	253	350	461
79	165	277	394	482

From each bottle three sub-samples were analysed using ICP-MS. 0.25 g of sample was weighed and dissolved in nitric acid. The homogeneity contribution to the total uncertainty was calculated using a 1-way-ANOVA (see Annex 1).

The estimates of analyte-specific inhomogeneity contributions u_{bb} to be included into the total uncertainty budgets were calculated according to ISO Guide 35 [4] using Eq. (1):

$$s_{bb} = \sqrt{\frac{MS_{\text{among}} - MS_{\text{within}}}{n}} \quad (1)$$

with:

- MS_{among} mean of squared deviations between bottles (from 1-way ANOVA)
- MS_{within} mean of squared deviations within bottles (from 1-way ANOVA)
- n number of replicate sub-samples per bottle

In cases, where $S_{\text{within}} > S_{\text{between}}$ the inhomogeneity contribution to the combined uncertainty is set to zero. This was the case for the elements Pb, Ag, Cd, Fe, Ni, Sb, Bi, Co, and V.

5. Stability testing

According to the producer of the material zinc powder should be stable because of the oxide layer on the surface of the particles. A possible instability could be the oxidation of Zn to ZnO.

To ensure the stability of the material the content of zinc oxide of the material was determined in February 2018 and in May 2019, using the following test:

- 1) 4 g of zinc powder was taken from the same bottle and transferred into a beaker filled with 20 ml of ammonia and 30 ml of water followed by strong mixing for 40 s.
- 2) The whole mixture was transferred immediately into a 100-ml-volumetric flask, filled up to the mark with water and mixed well.
- 3) 50 ml of this mixture were filtered, transferred into a 100-ml-volumetric flask filled with 10 ml of concentrated HCl.
- 4) The zinc-content of the solution was measured using ICP-OES

Table 3: Results of stability measurements

	Sample	1	2	3	Mean	s
02/2018	1	0.130	0.121	0.117	0.114	0.012
	2	0.112	0.103	0.101		
05/2019	1	0.097	0.139	0.120	0.117	0.017
	2	0.133	0.109	0.104		

The test results don't indicate an additional oxidation within 15 months of storage.

To consider a possible oxidation it was assumed that the mass fraction of ZnO increases from 0.1 % in 02/2018 to 0.5 % within ten years. This would increase the weight of the material by 0.1 %. As a result, the mass fractions of the certified impurities will decrease by 0.1 %. Therefore, an uncertainty contribution of 0.1 % is added to the combined uncertainty.

The stability test will be repeated regularly.

6. Characterisation study

6.1 Analytical methods

14 laboratories participated in the certification inter-laboratory comparison. All laboratories were highly experienced in the analysis of zinc or participated successfully in former certification inter-laboratory comparisons. For some elements part of the laboratories used more than one analytical method reporting more than one data set.

The laboratories were asked to analyse five subsamples. They were free to choose any suitable analytical method for their determinations. Table 4 shows the analytical methods used by the participating laboratories.

Table 4: Analytical procedures used by the participating laboratories

Lab-No.	Element.	Sample mass	Sample pretreatment	Analytical method
1	Ag, Al, As, Bi, Cd, Co, Cu, Fe, Ni, Pb, Sb, Sn, V	5 g	Dissolution with HCl/HNO ₃ ,	ICP-OES, calibration with commercial solutions (Carl Roth), standard addition
2	Pb, Cd, Fe, Cu, Ni, Sn	1 g	Dissolution with HNO ₃ /tartaric acid according to DIN ISO 3815-2	ICP-OES, calibration with commercial solutions
3	Pb, Cd, Fe, Cu, Ni, Sn, Sb, Ag, Al, As, Bi, Tl, V	5 g	Dissolution with HCl/HNO ₃ ,	ICP-OES, calibration with commercial solutions (Merck)
	Pb, Cd, Fe, Cu, Ni, Sn, Sb, Ag, Al, As, Bi, Tl, V	5 g	Dissolution with HCl/HNO ₃ ,	ICP-MS, calibration with commercial solutions (Merck)
4	Pb, Cd, Fe, Cu, Ni, Sn, Sb	2.5 g	Dissolution with HCl/HNO ₃ ,	ICP-OES, calibration with commercial solutions (Kraft)
5	Pb, Cd, Fe, Cu, Ni, Sn	2.5 g	Dissolution with HCl/HNO ₃	ICP-OES, calibration with commercial mono-element solutions
6	Pb, Cd, Fe, Cu, Ni, Sn	1 g	Dissolution with HCl/H ₂ O ₂	ICP-OES, matrix matched calibration with commercial mono-element solutions (traceable to NIST standards)
7	Al, Sb, As, Bi, Pb, Cd, Co, Fe, In, Cu, Ni, Ag, Tl, V, Sn	10 g	Dissolution with HNO ₃ , HCl	ICP-MS, calibration with commercial solutions (Merck)
8	Pb, Cd, Fe, Cu, Ni, Sn, Sb	1 g	Dissolution with HNO ₃	ICP-MS, matrix matched calibration with commercial mono-element solutions
	Pb, Cd, Fe, Cu, Ni	1 g	Dissolution with HNO ₃	ICP-OES, matrix matched calibration with commercial mono-element solutions
9	Pb, Cd, Fe, Cu, Ni, Sb	10 g	Dissolution with HNO ₃ , HCl	ICP-OES with matrix matched standards, prepared from pure metals
	Sn	5 g	Dissolution with HCl	ICP-OES with matrix matched standards, prepared from pure metal
10	Pb, Cd, Fe, Cu, Ni, Sn, Sb	2 g	Dissolution with HNO ₃	ICP-MS, calibration with matrix matched commercial solution standards
11	Pb, Cd, Fe, Cu, Ni, Tl	0.5 g	Dissolution with HNO ₃	ETAAS, calibration with commercial solutions (Merck)
	Pb, Ag, Cd, Cu, Fe, Ni, Sb, Tl, Bi, Co, V	0.25 g	Dissolution with HNO ₃	ICP-MS, calibration with commercial solutions (Merck)
12	Pb, Cd, Fe, Cu, Ni, Sn, Sb, Tl, Al, Bi, Co, V	2 g	Dissolution with HCl/HNO ₃	ICP-MS, calibration with matrix matched commercial solution standards
13	Pb, Cd, Fe, Cu, Ni, Sb	1 g	Dissolution with HNO ₃	ICP-OES, calibration with commercial mono-element solutions
14	Pb, Cd, Fe, Cu, Ni, Sn, Sb, Ag, Al, As, Bi, Co, In, Tl, V	2 g	Dissolution with HNO ₃	ICP-OES, calibration with commercial mono-element solutions

For all analytical methods where a calibration was necessary this was performed using liquid standard solutions. All participating laboratories were asked to use only standard solutions prepared from pure metals or stoichiometric compounds or internally checked commercial calibration solutions.

6.2 Analytical results and statistical evaluation

The analytical results of the inter-laboratory certification comparison are listed in Tables 5 to 19. These tables show the single results (M_i) of each laboratory, the respective laboratories' mean values (M), absolute and relative intra-laboratory standard deviation (s and s_{rel} , respectively), the standard deviation of laboratory means (s_M), and in addition the square root of mean of variances of data sets under repeatability conditions (\bar{s}_i). n is the number of accepted data sets. The continuous line marks the certified value (mean of the laboratories' means), the broken lines mark the standard deviation, calculated from the laboratories' means. Outliers which have been excluded are highlighted. In the related figures for each laboratory its mean value and single standard deviation is given.

Table 5: Results for Pb

Lab./Meth.	1_ICP	5_ICP	7_ICP-MS	11_ETAAS	11_ICP-MS	6_ICP	8_ICP-MS	14_ICP	13_ICP	10_ICP-MS	8_ICP	12_ICP-MS	3_ICP-MS	3_ICP	2_ICP	9_ICP	4_ICP		
M_i [mg/kg]	12.6	14	14.6	14.92	15.09	15	15.36	15.70	16.15	16.11	15.83	16.03	16.95	16.72	17	17.3	20		n 15
	12.7	15	14.6	14.75	15.06	16	15.82	15.77	15.52	16.02	15.68	16.12	16.24	16.71	17	17.3	20		
	12.7	14	14.4	14.90	14.95	15	15.51	15.75	15.56	15.95	15.88	16.02	15.92	16.87	17	17.2	20		
	12.7	14	14.7	14.39	15.09	16	15.83	15.75	15.83	15.72	15.96	16.03	17.49	16.76	16	17.3	19		
	12.8	14	14.4	14.75	15.07	16	15.94	15.76	16.03	15.61	16.07	16.09			17	17.2	20		
			14.98	15.02	15.08			15.78	15.82										
M [mg/kg]	12.68	14.20	14.54	14.78	15.05	15.60	15.69	15.75	15.82	15.88	15.88	16.06	16.65	16.76	16.80	17.26	19.80		15.78
s [mg/kg]	0.073	0.447	0.134	0.213	0.049	0.548	0.247	0.028	0.249	0.209	0.144	0.045	0.706	0.073	0.447	0.055	0.447	s_M [mg/kg]	0.872
s_{rel}	0.006	0.031	0.009	0.014	0.003	0.035	0.016	0.002	0.016	0.013	0.009	0.003	0.042	0.004	0.027	0.003	0.023	\bar{s}_i [mg/kg]	0.312
																			0.055




Table 6: Results for Ag

Lab./Meth.	1_ICP	3_ICP	11_ICP-MS	14_ICP	3_ICP-MS		
M_i [mg/kg]	0.84	1.01	1.10	1.06	1.12		n
	0.83	1.00	1.01	1.03	1.02		5
	0.85	1.01	1.02	1.02	1.03		
	0.84	1.01	1.00	1.03	1.11		
	0.85		1.00	1.04			
	0.84		0.97	1.11			
			1.02				
M [mg/kg]	0.84	1.01	1.02	1.05	1.07		1.00
s [mg/kg]	0.008	0.007	0.040	0.033	0.052	s_M [mg/kg]	0.090
						\bar{s}_i [mg/kg]	0.033
s_{rel}	0.009	0.007	0.039	0.032	0.049		0.091




Table 7: Results for Cd

Lab./Meth.	6_ICP	11_ICP-MS	1_ICP	9_ICP	13_ICP	8_ICP	10_ICP-MS	12_ICP-MS	8_ICP-MS	14_ICP	3_ICP	3_ICP-MS	7_ICP-MS	11_ETAAS	5_ICP	4_ICP		
M_i [mg/kg]	0.8	1.27	1.34	1.4	1.67	1.62	1.67	1.71	1.71	1.74	1.76	1.86	1.8	1.96	2	2		n
	0.8	1.27	1.34	1.4	1.59	1.66	1.64	1.70	1.68	1.75	1.75	1.66	1.8	1.94	2	2		15
	0.8	1.28	1.35	1.4	1.61	1.62	1.65	1.69	1.73	1.74	1.77	1.69	1.8	2.04	2	2		
	0.8	1.29	1.34	1.4	1.63	1.66	1.64	1.69	1.71	1.74	1.77	1.86	1.8	1.95	2	2		
	0.8	1.26	1.36	1.4	1.64	1.66	1.68	1.70	1.73	1.75	1.75		1.8	1.97		2		
		1.28	1.35		1.59					1.75				2.10				
M [mg/kg]	0.80	1.28	1.35	1.40	1.62	1.64	1.65	1.70	1.71	1.75	1.76	1.77	1.80	1.99	2.00	2.00		1.69
s [mg/kg]	0.000	0.009	0.008	0.000	0.031	0.023	0.016	0.010	0.019	0.004	0.009	0.108	0.000	0.063	0.000	0.000	s_M [mg/kg]	0.222
s_{rel}	0.000	0.007	0.006	0.000	0.019	0.014	0.010	0.006	0.011	0.002	0.005	0.061	0.000	0.032	0.000	0.000	\bar{s}_i [mg/kg]	0.035
																		0.131




Table 8: Results for Cu

Lab./Meth.	1_ICP	4_ICP	7_ICP-MS	10_ICP-MS	11_ICP-MS	13_ICP	3_ICP	14_ICP	12_ICP-MS	8_ICP	8_ICP-MS	5_ICP	11_ETAAS	3_ICP-MS	9_ICP		
M_i [mg/kg]	3.0	3	3.1	3.06	3.69	3.70	3.68	3.82	3.88	3.86	4.02	4	4.28	4.23	3.8		n 15
	2.9	3	3.3	3.15	3.64	3.64	3.67	3.83	3.86	4.03	3.99	4	4.08	3.97	3.8		
	3.0	3	3.2	3.53	3.61	3.61	3.73	3.81	3.87	4.09	3.94	4	3.86	3.84	3.7		
	2.9	3	3.2	3.47	3.62	3.71	3.74	3.83	3.89	3.92	4.04	4	4.18	4.21	4.4		
	3.0	3	3.2	3.45	3.59	3.63		3.83	3.81	3.93	3.93	4	3.92		5.1		
	3.0				3.62	3.61			3.95					4.05			
					3.60												
M [mg/kg]	2.96	3.00	3.20	3.33	3.62	3.65	3.71	3.84	3.86	3.97	3.98	4.00	4.06	4.07	4.16		3.69
s [mg/kg]	0.022	0.000	0.071	0.212	0.030	0.044	0.037	0.052	0.031	0.093	0.047	0.000	0.157	0.191	0.594	s_M [mg/kg]	0.396
s_{rel}	0.007	0.000	0.022	0.064	0.008	0.012	0.010	0.013	0.008	0.023	0.012	0.000	0.039	0.047	0.143	\bar{s}_i [mg/kg]	0.179
																	0.107




Table 9: Results for Fe

Lab./Meth.	2_ICP	9_ICP	4_ICP	8_ICP-MS	14_ICP	7_ICP-MS	8_ICP	12_ICP-MS	1_ICP	11_ICP-MS	13_ICP	11_ETAAS	3_ICP	3_ICP-MS	10_ICP-MS	5_ICP	6_ICP		
M_i [mg/kg]	1	1.9	2	2.02	2.01	2.0	1.98	2.07	2.80	2.44	2.19	2.59	2.61	2.67	2.63	3	4		n 14
	1	1.8	2	2.06	2.01	2.1	2.11	2.07	2.78	1.77	2.20	2.58	2.40	2.36	2.45	4	3		
	1	1.6	2	1.84	2.02	2.0	2.05	2.07	1.76	2.16	2.24	2.49	2.62	2.52	2.58	2	3		
	1	1.7	2	2.13	2.02	2.0	2.10	2.06	1.73	2.23	2.18	2.59	2.42	2.52		2	3		
	1	1.7 1.4	2	2.05	2.02 2.03	2.0	2.04	2.06	1.87 1.91	2.00 2.16	2.22 2.24	2.38 2.39				2			
M [mg/kg]	1.00	1.68	2.00	2.02	2.02	2.02	2.06	2.07	2.14	2.20	2.21	2.50	2.51	2.52	2.56	2.60	3.25		2.18
s [mg/kg]	0.000	0.189	0.000	0.108	0.009	0.045	0.054	0.004	0.507	0.278	0.026	0.097	0.118	0.124	0.091	0.894	0.500	s_M [mg/kg]	0.258
s_{rel}	0.000	0.113	0.000	0.054	0.004	0.022	0.026	0.002	0.237	0.127	0.012	0.039	0.047	0.049	0.036	0.344	0.154	\bar{s}_i [mg/kg]	0.176
																			0.119




Table 10: Results for Ni

Lab./Meth.	7_ICP-MS	9_ICP	10_ICP-MS	13_ICP	8_ICP-MS	8_ICP	12_ICP-MS	11_ICP-MS	1_ICP	14_ICP	3_ICP-MS	3_ICP	11_ETAAS		
M_i [mg/kg]	0.31	0.31	0.50	0.36	0.39	0.36	0.40	0.48	0.42	0.51	0.58	0.57	0.56		n
	0.31	0.36	0.41	0.40	0.41	0.39	0.41	0.42	0.48	0.52	0.53	0.59	0.58		13
	0.30	0.32	0.31	0.41	0.36	0.40	0.41	0.33	0.45	0.53	0.49	0.54	0.59		
	0.29	0.34	0.35	0.40	0.41	0.44	0.40	0.49	0.39	0.52	0.48	0.49	0.55		
	0.28	0.34	0.36	0.39	0.40	0.41	0.39	0.32	0.42	0.51			0.52		
				0.36				0.34	0.40	0.54					
								0.44							
M [mg/kg]	0.30	0.33	0.38	0.39	0.39	0.40	0.40	0.40	0.43	0.52	0.52	0.55	0.56		0.43
s [mg/kg]	0.013	0.019	0.077	0.022	0.019	0.027	0.008	0.073	0.033	0.010	0.046	0.041	0.026	s_M [mg/kg]	0.082
s_{rel}	0.044	0.058	0.199	0.056	0.048	0.067	0.019	0.182	0.078	0.019	0.087	0.075	0.047	\bar{s}_i [mg/kg]	0.038
															0.191




Table 11: Results for Sb

Lab./Meth.	7_ICP-MS	11_ICP-MS	12_ICP-MS	3_ICP-MS	8_ICP-MS	13_ICP	9_ICP	14_ICP	6_ICP	10_ICP-MS	1_ICP		
M_i [mg/kg]	0.02	0.02	0.04	0.06	0.07	0.18	0.19	1.41	2	< 0.3	< 0.5		n
	0.01	0.02	0.04	0.05	0.05	0.03	0.14	1.07	1	< 0.3	< 0.5		5
		0.04	0.04	0.05	0.05	0.23	0.22	0.98	1	< 0.3	< 0.5		
		0.03	0.04	0.05	0.06	-0.06	0.28	1.52	1	< 0.3	< 0.5		
		0.07	0.04		0.06	0.48	0.22	1.30	1		< 0.5		
	0.03					0.21		1.42					
	0.04												
M [mg/kg]	0.02	0.04	0.04	0.05	0.06	0.18	0.21	1.28	1.20	< 0.3	< 0.5		0.040
s [mg/kg]	0.007	0.016	0.001	0.003	0.006	0.186	0.051	0.211	0.447			s_M [mg/kg]	0.016
s_{rel}	0.471	0.447	0.014	0.048	0.106	1.045	0.243	0.165	0.373			\bar{s}_i [mg/kg]	0.008
													0.407




Table 12: Results for TI

Lab./Meth.	7_ICP-MS	11_ICP-MS	3_ICP	3_ICP-MS	12_ICP-MS	11_ETAAS	14_ICP		
M_i [mg/kg]	3.5	3.65	3.73	3.82	3.82	3.84	4.31		n
	3.5	3.58	3.74	3.69	3.83	4.01	4.35		7
	3.4	3.60	3.81	3.66	3.83	3.95	4.30		
	3.5	3.59	3.83	4.03	3.78	3.71	4.37		
	3.4	3.57			3.83	3.98	4.37		
		3.63				3.95	4.31		
	3.57								
M [mg/kg]	3.46	3.60	3.78	3.80	3.82	3.91	4.34		3.81
s [mg/kg]	0.055	0.032	0.052	0.166	0.021	0.112	0.032	s_M [mg/kg]	0.276
								\bar{s}_i [mg/kg]	0.083
s_{rel}	0.016	0.009	0.014	0.044	0.006	0.029	0.007		0.072




Table 13: Results for Al

Lab./Meth.	12_ICP-MS	1_ICP	7_ICP-MS	14_ICP	3_ICP		
M_i [mg/kg]	< 0.15	0.200	0.140	0.219	< 0.25		n
	< 0.15	0.200	0.130	0.293	< 0.25		3
	< 0.15	0.180	0.160	0.336	< 0.25		
	< 0.15	0.170	0.180	0.441	< 0.25		
	< 0.15	0.260	0.240	0.402	< 0.25		
		0.130		0.288			
		0.008					
M [mg/kg]	< 0.15	0.164	0.170	0.330	< 0.25		0.221
s [mg/kg]		0.079	0.044	0.081		s_M [mg/kg]	0.094
						\bar{s}_i [mg/kg]	0.070
s_{rel}		0.482	0.256	0.246			0.424




Table 14: Results for As

Lab./Meth.	7_ICP-MS	3_ICP-MS	14_ICP	1_ICP	1_ICP	3_ICP			
M_i [mg/kg]	< 0.01	< 0.02	0.128	0.200	< 0.5	< 1			n
	< 0.01	< 0.02	0.128	0.200	< 0.5	< 1			1
	< 0.01	< 0.02	0.154	0.180	< 0.5	< 1			
	< 0.01	< 0.02	0.136	0.170	< 0.5	< 1			
	< 0.01	< 0.02	0.107	0.260	< 0.5	< 1			
			0.160	0.130					
				0.008					
M [mg/kg]	< 0.01	< 0.02	0.135	0.164	< 0.5	< 1			0.150
s [mg/kg]			0.019	0.079				s_M [mg/kg]	0.020
								\bar{s}_i [mg/kg]	0.014
s_{rel}			0.143	0.482					0.135

Table 15: Results for In

Lab./Meth.	14_ICP	7_ICP-MS	3_ICP-OES		
M_i [mg/kg]	0.113	0.50	< 0.2		n
	0.124	0.40	< 0.2		2
	0.085	0.40	< 0.2		
	0.115	0.40	< 0.2		
	0.113	0.50			
	0.137				
M [mg/kg]	0.114	0.440	< 0.2		0.277
s [mg/kg]	0.017	0.055		s_M [mg/kg]	0.230
				\bar{s}_i [mg/kg]	0.041
s_{rel}	0.150	0.124			0.831

Table 16: Results for Bi

Lab./Meth.	12_ICP-MS	11_ICP-MS	3_ICP-MS	1_ICP	14_ICP		
M_i [mg/kg]	0.093	0.111	0.109	< 0.2	0.541		n
	0.094	0.103	0.099	< 0.2	0.529		3
	0.128	0.105	0.099	< 0.2	0.548		
	0.093	0.100	0.104	< 0.2	0.417		
	0.100	0.103		< 0.2	0.488		
		0.089	0.108		< 0.2	0.538	
M [mg/kg]	0.102	0.103	0.103	< 0.2	0.510		0.102
s [mg/kg]	0.015	0.007	0.005		0.051	s_M [mg/kg]	0.001
						\bar{s}_i [mg/kg]	0.010
s_{rel}	0.150	0.068	0.050		0.099		0.006




Table 17 Results for Co

Lab./Meth.	7_ICP-MS	11_ICP-MS	12_ICP-MS	14_ICP	1_ICP	3_ICP		
M_i [mg/kg]	0.030	0.041	0.044	0.117	< 0.2	<0.2		n
	0.040	0.041	0.044	0.125	< 0.2	<0.2		3
	0.040	0.040	0.042	0.111	< 0.2	<0.2		
	0.040	0.040	0.043	0.108	< 0.2	<0.2		
	0.040	0.040	0.043	0.118	< 0.2	< 0.2		
		0.040	0.041	0.113	< 0.2	< 0.2		
M [mg/kg]	0.038	0.040	0.043	0.115	< 0.2	<0.2		0.041
s [mg/kg]	0.004	0.001	0.001	0.006			s_M [mg/kg]	0.003
							\bar{s}_i [mg/kg]	0.004
s_{rel}	0.118	0.013	0.017	0.052				0.064




Table 18: Results for Sn

Lab./Meth.	8_ICP-MS	7_ICP-MS	14_ICP	9_ICP	10_ICP-MS	1_ICP	6_ICP		
M_i [mg/kg]	0.022	0.03	0.34	1.24	< 0.3	< 1	1		n 4
	0.027	0.03	0.42	1.15	< 0.3	< 1	1		
	0.016	0.03	0.92	1.08	< 0.3	< 1	< 1		
	0.037	0.03	0.37	1.16	< 0.3	< 1	< 1		
	0.021	0.03	0.40	1.33	< 1	< 1			
		0.64		< 1					
M [mg/kg]	0.025	0.030	0.515	1.192	< 0.3	< 1	< 1		0.440
s [mg/kg]	0.008	0.000	0.224	0.096				s_M [mg/kg]	0.551
s_{rel}	0.324	0.000	0.434	0.080				\bar{s}_i [mg/kg]	0.099
									1.252





Table 19: Results for V

Lab./Meth.	11_ICP-MS	14_ICP	7_ICP-MS	3_ICP-MS	3_ICP-OES		
M_i [mg/kg]	0.008	0.034	0.110	< 0.02	< 0.2		n
	0.008	0.032	0.120	< 0.02	< 0.2		3
	0.016	0.032	0.120	< 0.02	< 0.2		
	0.009	0.041	0.110	< 0.02	< 0.2		
	0.012	0.033	0.080				
	0.010	0.033					
	0.008						
M [mg/kg]	0.010	0.034	0.108	< 0.02	< 0.2		0.051
s [mg/kg]	0.003	0.004	0.016			s_M [mg/kg]	0.051
s_{rel}	0.273	0.104	0.152			\bar{s}_i [mg/kg]	0.010
							1.004



The statistical evaluation of the data was performed using the software program SoftCRM 1.2.2. [6]. The following results were obtained:

Tab. 20: Outcome of statistical tests on the results obtained for Pb

	1 st run	2 nd run
Number of data sets	16	14
Scheffe's test (data compatible?)	yes	yes
Snedecor-F-Test and Bartlett-Test	Pooling not allowed	Pooling not allowed
Dixon ($\alpha = 0.05$)	Lab. 4	---
Dixon ($\alpha = 0.01$)	---	---
Nalimov ($\alpha = 0.05$)	Lab. 4, Lab. 1	---
Nalimov ($\alpha = 0.01$)	Lab. 4	---
Grubbs ($\alpha = 0.05$)	Lab. 4	---
Grubbs ($\alpha = 0.01$)	---	---
Grubbs Pair ($\alpha = 0.05$)	---	---
Grubbs Pair ($\alpha = 0.01$)	---	---
Cochran	---	---
Kolmogorov-Smirnov-Lilliefors Test ($\alpha = 0.05$)	Distribution: normal	Distribution: normal
Kolmogorov-Smirnov-Lilliefors Test ($\alpha = 0.01$)	Distribution: not normal	Distribution: normal
Skewness & Kurtosis Test ($\alpha = 0.05$)	Distribution: normal	Distribution: normal
Skewness & Kurtosis Test ($\alpha = 0.01$)	Distribution: normal	Distribution: normal

The outliers (Lab. 4 and Lab. 1) were removed.

Tab. 21: Outcome of statistical tests on the results obtained for Ag and Cu

	Ag	Cu
Number of data sets	5	15
Scheffe's test (data compatible?)	yes	yes
Snedecor-F-Test and Bartlett-Test	Pooling not allowed	Pooling not allowed
Dixon ($\alpha = 0.05$)	Lab. 1	---
Dixon ($\alpha = 0.01$)	---	---
Nalimov ($\alpha = 0.05$)	Lab. 1	---
Nalimov ($\alpha = 0.01$)	Lab. 1	---
Grubbs ($\alpha = 0.05$)	Lab. 1	---
Grubbs ($\alpha = 0.01$)	---	---
Grubbs Pair ($\alpha = 0.05$)	---	---
Grubbs Pair ($\alpha = 0.01$)	---	---
Cochran	---	---
Kolmogorov-Smirnov-Lilliefors Test ($\alpha = 0.05$)	Distribution: not normal	Distribution: normal
Kolmogorov-Smirnov-Lilliefors Test ($\alpha = 0.01$)	Distribution: normal	Distribution: normal
Skewness & Kurtosis Test ($\alpha = 0.05$)	Distribution: insuff. data	Distribution: normal
Skewness & Kurtosis Test ($\alpha = 0.01$)	Distribution: insuff. data	Distribution: normal

The outlier (Lab. 1; Ag) was not removed.

Tab. 22: Outcome of statistical tests on the results obtained for Cd

	1 st run	2 nd run
Number of data sets	16	15
Scheffe's test (data compatible?)	yes	yes
Snedecor-F-Test and Bartlett-Test	Pooling not allowed	Pooling not allowed
Dixon ($\alpha = 0.05$)	---	---
Dixon ($\alpha = 0.01$)	---	---
Nalimov ($\alpha = 0.05$)	Lab. 6	Lab. 11_ICP-MS
Nalimov ($\alpha = 0.01$)	Lab. 6	---
Grubbs ($\alpha = 0.05$)	Lab. 6	---
Grubbs ($\alpha = 0.01$)	---	---
Grubbs Pair ($\alpha = 0.05$)	---	---
Grubbs Pair ($\alpha = 0.01$)	---	---
Cochran	---	---
Kolmogorov-Smirnov-Lilliefors Test ($\alpha = 0.05$)	Distribution: not normal	Distribution: normal
Kolmogorov-Smirnov-Lilliefors Test ($\alpha = 0.01$)	Distribution: normal	Distribution: normal
Skewness & Kurtosis Test ($\alpha = 0.05$)	Distribution: not normal	Distribution: normal
Skewness & Kurtosis Test ($\alpha = 0.01$)	Distribution: normal	Distribution: normal

The outlier (Lab. 6) was removed, the straggler (Lab. 11) was not removed.

Tab. 23: Outcome of statistical tests on the results obtained for Fe

	1 st run	2 nd run
Number of data sets	17	14
Scheffe's test (data compatible?)	yes	yes
Snedecor-F-Test and Bartlett-Test	Pooling not allowed	Pooling not allowed
Dixon ($\alpha = 0.05$)	Lab. 2, Lab. 6	---
Dixon ($\alpha = 0.01$)	Lab. 2	---
Nalimov ($\alpha = 0.05$)	Lab. 2, Lab. 6	Lab. 9
Nalimov ($\alpha = 0.01$)	Lab. 2	---
Grubbs ($\alpha = 0.05$)	Lab. 2	---
Grubbs ($\alpha = 0.01$)	---	---
Grubbs Pair ($\alpha = 0.05$)	---	---
Grubbs Pair ($\alpha = 0.01$)	---	---
Cochran	Lab. 5	---
Kolmogorov-Smirnov-Lilliefors Test ($\alpha = 0.05$)	Distribution: not normal	Distribution: normal
Kolmogorov-Smirnov-Lilliefors Test ($\alpha = 0.01$)	Distribution: normal	Distribution: normal
Skewness & Kurtosis Test ($\alpha = 0.05$)	Distribution: not normal	Distribution: normal
Skewness & Kurtosis Test ($\alpha = 0.01$)	Distribution: normal	Distribution: normal

The outliers (Labs. 5, 2 and 6, 1st run) were removed, the straggler (Lab. 9, 2nd run) was not removed.

Tab. 24: Outcome of statistical tests on the results obtained for Ni and Sb (without all ICP-OES results since this method was not suitable to analyse this element)

	Ni	Sb
Number of data sets	13	5
Scheffe's test (data compatible?)	yes	yes
Snedecor-F-Test and Bartlett-Test	Pooling not allowed	Pooling not allowed
Dixon ($\alpha = 0.05$)	---	---
Dixon ($\alpha = 0.01$)	---	---
Nalimov ($\alpha = 0.05$)	---	---
Nalimov ($\alpha = 0.01$)	---	---
Grubbs ($\alpha = 0.05$)	---	---
Grubbs ($\alpha = 0.01$)	---	---
Grubbs Pair ($\alpha = 0.05$)	---	---
Grubbs Pair ($\alpha = 0.01$)	---	---
Cochran	---	---
Kolmogorov-Smirnov-Lilliefors Test ($\alpha = 0.05$)	Distribution: normal	Distribution: normal
Kolmogorov-Smirnov-Lilliefors Test ($\alpha = 0.01$)	Distribution: normal	Distribution: normal
Skewness & Kurtosis Test ($\alpha = 0.05$)	Distribution: normal	Insufficient data
Skewness & Kurtosis Test ($\alpha = 0.01$)	Distribution: normal	Insufficient data

Tab. 25: Outcome of statistical tests on the results obtained for Tl

	Tl
Number of data sets	7
Scheffe's test (data compatible?)	yes
Snedecor-F-Test and Bartlett-Test	Pooling not allowed
Dixon ($\alpha = 0.05$)	---
Dixon ($\alpha = 0.01$)	---
Nalimov ($\alpha = 0.05$)	Lab. 14
Nalimov ($\alpha = 0.01$)	---
Grubbs ($\alpha = 0.05$)	---
Grubbs ($\alpha = 0.01$)	---
Grubbs Pair ($\alpha = 0.05$)	---
Grubbs Pair ($\alpha = 0.01$)	---
Cochran	---
Kolmogorov-Smirnov-Lilliefors Test ($\alpha = 0.05$)	Distribution: normal
Kolmogorov-Smirnov-Lilliefors Test ($\alpha = 0.01$)	Distribution: normal
Skewness & Kurtosis Test ($\alpha = 0.05$)	Distribution: normal
Skewness & Kurtosis Test ($\alpha = 0.01$)	Distribution: normal

The straggler (Lab. 14, Tl) was not removed.

For the other elements no statistical evaluation using SoftCRM was performed because the number of data sets was too low.

The resp. combined uncertainties were calculated from the spread resulting from the certification inter-laboratory comparison (u_{ilc}) and the uncertainty contribution from possible inhomogeneity of the material using Equation 2.

$$u_{comb} = \sqrt{u_{ilc}^2 + u_{bb}^2 + u_{stab}^2} \quad (2)$$

with

$$u_{ilc} = \sqrt{\frac{s_M^2}{n}} : \text{uncertainty contribution resulting from inter-laboratory comparison}$$

n : number of data sets used for calculating the certified mass fraction of each element

Table 26: Uncertainty calculation

	uncertainty contribution from						u (comb)	U	uncertainty contribution from		
	M	n	s_M	u_{ilc}	u_{bb}	u_{stab}			homogeneity test (rel. %)		
	mg/kg		mg/kg	mg/kg	mg/kg		mg/kg	mg/kg			
Pb	15.780	15	0.8720	0.2251	0.00000	0.01578	0.2257	0.4514		0.0000	
Ag	1.000	5	0.0900	0.0402	0.00000	0.00100	0.0403	0.0805		0.0000	
Cd	1.690	15	0.2220	0.0573	0.00000	0.00169	0.0573	0.1147		0.0000	
Cu	3.690	15	0.3960	0.1022	0.02013	0.00369	0.1043	0.2085		0.5454	
Fe	2.180	14	0.2580	0.0690	0.00000	0.00218	0.0690	0.1380		0.0000	
Ni	0.430	13	0.0820	0.0227	0.00000	0.00043	0.0227	0.0455		0.0000	
Sb	0.040	6	0.0160	0.0065	0.00000	0.00004	0.0065	0.0131		0.0000	
Tl	3.810	7	0.2760	0.1043	0.03543	0.00381	0.1102	0.2205		0.9300	
Al	0.221	3	0.0940	0.0543	0.04738	0.00022	0.0720	0.1441		21.4401	
Bi	0.102	3	0.0010	0.0006	0.00000	0.00010	0.0006	0.0012		0.0000	
Co	0.041	3	0.0030	0.0017	0.00000	0.00004	0.0017	0.0035		0.0000	

The expanded uncertainties U are calculated by multiplication of $u_{combined}$ with a coverage factor of $k = 2$ using Equation 3.

$$U = k \cdot u_{combined} \quad (3)$$

The calculated mass fractions and their resp. expanded uncertainties are given on Page 3 of this report. Rounding was done according to DIN 1333 [7].

7. Instructions for users

The certified reference material BAM-M603 is intended for the development, validation and quality control of analytical methods and procedures for the determination of trace components in pure zinc. For wet chemical analysis, a minimum sample intake of 0.25 g should be used.

8. Metrological Traceability

To ensure traceability of the certified mass fractions to the SI (Système International d'Unités) calibration was done using standard solutions prepared from pure metals or stoichiometric compounds or well checked commercial calibration solutions.

9. Information on and purchase of the CRM

Certified reference material BAM-M603 is supplied by
Bundesanstalt für Materialforschung und -prüfung (BAM)
Division 1.6 „Inorganic Reference Materials“
Richard-Willstätter-Str. 11, D-12489 Berlin, Germany
Phone +49 (0)30 - 8104 2061
Fax: +49 (0)30 - 8104 72061
E-Mail: sales.crm@bam.de

Each bottle of BAM-M603 will be distributed together with a detailed certificate containing the certified values and their uncertainties, the mean values and standard deviations of all accepted data sets and information on the analytical methods used and the names of the participating laboratories. Information on certified reference materials can be obtained from BAM:
<https://www.bam.de>.
Tel. +49 30 8104 1111.

10. References

- [1] DIN EN 1179 Zinc and zinc alloys - Primary zinc; German version EN 1179:2003
- [2] DIN EN ISO 17034, Allgemeine Anforderungen an die Kompetenz von Referenzmaterialherstellern, 2016
- [3] ISO Guide 31, Reference materials - Contents of certificates, labels and accompanying documentation, 2015
- [4] ISO Guide 35, Reference materials - Guidance for characterization and assessment of homogeneity and stability, 2017
- [5] ASTM B214 - 16, Standard Test Method for Sieve Analysis of Metal Powders, ASTM International, West Conshohocken, PA, 2016
- [6] Bonas G, Zervou M, Papaeoannou T, Lees M: Accred Qual Assur (2003) 8:101-107
- [7] DIN 1333:1992-02 Zahlenangaben

Silver:

	1	2	3									
603-1	1.04266461	0.96210442	1.59209694			<i>Bottle</i>	<i>Number</i>	<i>Sum</i>	<i>Mean</i>	<i>Variance</i>		
603-2	1.03155999	0.95564134	1.00669165			1	3	3.59686597	1.19895532	0.11754273		
603-3	1.03702951	0.99711638	0.96718615			2	3	2.99389298	0.99796433	0.00149803		
603-4	1.02098036	0.98573506	1.03482048			3	3	3.00133204	1.00044401	0.00122783		
603-5	1.08136954	0.99068609	0.9772048			4	3	3.0415359	1.0138453	0.00064053		
603-6	1.15472168	0.98270338	0.96017954			5	3	3.04926044	1.01642015	0.00320925		
603-7	1.060899	0.97285992	0.95657698			6	3	3.0976046	1.03253487	0.01132404		
603-8	1.07661941	0.98520021	0.96182086			7	3	2.9903359	0.99677863	0.00314985		
603-9	0.96896244	0.9520668	0.95697984			8	3	3.02364049	1.00788016	0.00368046		
603-10	1.1851405	0.96965478	0.93877289			9	3	2.87800908	0.95933636	7.5531E-05		
603-11	0.97384132	0.97089113	0.96985188			10	3	3.09356816	1.03118939	0.01801413		
603-12	0.99004367	0.95670336	0.97560934			11	3	2.91458434	0.97152811	4.2832E-06		
603-13	1.19899356	1.0035938	0.96427194			12	3	2.92235638	0.97411879	0.00027956		
603-14	1.02345529	0.9611087	0.97714863			13	3	3.16685929	1.05561976	0.01580359		
603-15	0.99981418	1.11887901	0.95133561			14	3	2.96171262	0.98723754	0.00104811		
						15	3	3.07002879	1.02334293	0.0074329		
									1.01781304			
						ANOVA						
						<i>Source of variation</i>	<i>sums of squares (SS)</i>	<i>degrees of freedom (df)</i>	<i>Mean squares (MS)</i>	<i>F-value</i>	<i>P-value</i>	<i>critical F-value</i>
						Between groups	0.13298528	14	0.00949895	0.77047303	0.69060901	2.03742044
						Within groups	0.36986168	30	0.01232872			
						Total	0.50284696	44				
						within-sd	0.111035			status:	homogeneous	
						effective n	4.00					
						s_bb	0					
						s_bb_min	0.02821					
						u_bb	0.02821					
						Contribution to combined uncertainty:		0				

Copper:

	1	2	3									
603-1	3.6635424	3.6275236	3.68647797			<i>Bottle</i>	<i>Number</i>	<i>Sum</i>	<i>Mean</i>	<i>Variance</i>		
603-2	3.70341948	3.64520928	3.78524948			1	3	10.977544	3.65918132	0.00088317		
603-3	3.6770794	3.68372108	3.67660609			2	3	11.1338782	3.71129274	0.00494931		
603-4	3.54212941	3.53182524	3.69993175			3	3	11.0374066	3.67913552	1.5826E-05		
603-5	3.6446718	3.60720539	3.5977587			4	3	10.7738864	3.59129546	0.00887793		
603-6	3.59851265	3.60407407	3.62920491			5	3	10.8496359	3.6165453	0.00061564		
603-7	3.64540539	3.56072369	3.67869554			6	3	10.8317916	3.61059721	0.00026742		
603-8	3.61951986	3.6409348	3.59374386			7	3	10.8848246	3.62827488	0.00369943		
603-9	3.58974757	3.51916813	3.67226065			8	3	10.8541985	3.61806617	0.00055833		
603-10	3.61024276	3.57317246	3.5842284			9	3	10.7811763	3.59372545	0.0058712		
603-11	3.60700885	3.52577675	3.66375656			10	3	10.7676436	3.58921454	0.0003622		
603-12	3.69008788	3.60185293	3.61449178			11	3	10.7965422	3.59884739	0.00480956		
603-13	3.6002119	3.59201347	3.5198452			12	3	10.9064326	3.63547753	0.00227665		
603-14	3.6659176	3.61740769	3.57136199			13	3	10.7120706	3.57069019	0.00195571		
603-15	3.60542703	3.62995741	3.63834986			14	3	10.8546873	3.61822909	0.0022357		
						15	3	10.8737343	3.6245781	0.00029268		
									3.62301006			
						ANOVA						
						<i>Source of variation</i>	<i>sums of squares (SS)</i>	<i>degrees of freedom (df)</i>	<i>Mean squares (MS)</i>	<i>F-value</i>	<i>P-value</i>	<i>critical F-value</i>
						Between groups	0.05702338	14	0.0040731	1.62185466	0.13005577	2.03742044
						Within groups	0.07534149	30	0.00251138			
						Total	0.13236487	44				
						within-sd	0.050114			status:	homogeneous	
						effective n	4.00					
						s_bb	0.019759					
						s_bb_min	0.012732					
						u_bb	0.019759					
						Contribution to combined uncertainty [rel. %]:		0.54538273				

Iron:

	1	2	3									
603-1	2.64728927	3.16747897	1.52499725			<i>Bottle</i>	<i>Number</i>	<i>Sum</i>	<i>Mean</i>	<i>Variance</i>		
603-2	3.07407565	2.19608939	2.01280848			1	3	7.3397655	2.4465885	0.70464715		
603-3		1.77912541	1.71507126			2	3	7.28297352	2.42765784	0.32178996		
603-4	2.29120233	1.41428451	1.63707023			3	2	3.49419666	1.74709833	0.00205147		
603-5		1.48026727	2.89260448			4	3	5.34255706	1.78085235	0.20775119		
603-6		2.0284794	2.55707603			5	2	4.37287176	2.18643588	0.99734819		
603-7	1.72046867	3.92211322	1.85855375			6	2	4.58555543	2.29277772	0.1397072		
603-8	2.04316297		1.59291156			7	3	7.50113564	2.50037855	1.52076398		
603-9	2.5342946	1.25804648	1.62418718			8	2	3.63607452	1.81803726	0.10136317		
603-10	2.07640924	1.92699254	2.60482463			9	3	5.41652826	1.80550942	0.43186063		
603-11	1.53645273	2.32609578	1.52604024			10	3	6.60822641	2.20274214	0.12683408		
603-12	2.82554786	1.61770082	3.1150766			11	3	5.38858875	1.79619625	0.21062224		
603-13	3.4628893	3.57304064	1.88339552			12	3	7.55832528	2.51944176	0.63080926		
603-14	1.83752653		1.98820922			13	3	8.91932547	2.97310849	0.89363909		
603-15	2.20218557	2.9746597	3.04505701			14	2	3.82573575	1.91286788	0.01135264		
						15	3	8.22190227	2.74063409	0.21868405		
									2.21002176			
						ANOVA						
						<i>Source of variation</i>	<i>sums of squares (SS)</i>	<i>degrees of freedom (df)</i>	<i>Mean squares (MS)</i>	<i>F-value</i>	<i>P-value</i>	<i>critical F-value</i>
						Between groups	5.89649726	14	0.42117838	0.89333957	0.5752499	2.11110505
						Within groups	11.786626	25	0.47146504			
						Total	17.6831232	39				
						within-sd	0.686633			status:	homogeneous	
						effective n	4.00					
						s_bb	0					
						s_bb_min	0.182586					
						u_bb	0.182586					
						Contribution to combined uncertainty:	0					

Nickel:

	1	2	3									
603-1	0.45914851	0.58305152	0.24533115			Bottle	Number	Sum	Mean	Variance		
603-2	0.69973285	0.45725524	0.45527261			1	3	1.28753119	0.42917706	0.02918748		
603-3		0.29451233	0.68678487			2	3	1.6122607	0.53742023	0.01976002		
603-4	0.546732	0.28410759	0.30649407			3	2	0.9812972	0.4906486	0.07693887		
603-5		0.27589638	0.44598447			4	3	1.13733366	0.37911122	0.02119783		
603-6	0.3137757	0.31652465	0.29863853			5	2	0.72188084	0.36094042	0.01446498		
603-7	0.38197806	0.99082096	0.27903061			6	3	0.92893888	0.30964629	9.2767E-05		
603-8	0.50488823		0.28221518			7	3	1.65182963	0.55060988	0.1479889		
603-9	0.40821904	0.23451034	0.28130699			8	2	0.78710341	0.3935517	0.02479164		
603-10	0.3474397	0.30828216	0.33838241			9	3	0.92403638	0.30801213	0.00807855		
603-11	0.27327273	0.26902244	0.24513943			10	3	0.99410428	0.33136809	0.00042023		
603-12	0.44009191	0.27752831	0.51383547			11	3	0.7874346	0.2624782	0.00022999		
603-13	0.82640152	0.30559651	0.37905877			12	3	1.23145569	0.41048523	0.01461768		
603-14	0.31111761		0.3655607			13	3	1.5110568	0.5036856	0.07945835		
603-15	0.34543364	0.51973651	0.47096431			14	2	0.67667831	0.33833916	0.00148202		
						15	3	1.33613447	0.44537816	0.00808636		
									0.40339013			
						ANOVA						
						Source of variation	sums of squares (SS)	degrees of freedom (df)	Mean squares (MS)	F-value	P-value	critical F-value
						Between groups	0.31454547	14	0.02246753	0.75286179	0.70584729	2.09394853
						Within groups	0.77591384	26	0.02984284			
						Total	1.09045931	40				
						within-sd	0.172751			status:	homogeneous	
						effective n	4.00					
						s_bb	0					
						s_bb_min	0.045489					
						u_bb	0.045489					
						Contribution to combined uncertainty:	0					

Bismuth:

	1	2	3										
						<i>Bottle</i>	<i>Number</i>	<i>Sum</i>	<i>Mean</i>	<i>Variance</i>			
603-1	0.0957987	0.09531659	0.17672275			1	3	0.36783803	0.12261268	0.00219598			
603-2	0.14373142	0.07745933	0.07872286			2	3	0.29991361	0.0999712	0.00143662			
603-3	0.07249139	0.07925708	0.10744127			3	3	0.25918975	0.08639658	0.0003436			
603-4	0.14551801	0.07376878	0.1406387			4	3	0.35992549	0.11997516	0.00160722			
603-5	0.14349551	0.11018905	0.07701115			5	3	0.33069571	0.1102319	0.00110504			
603-6	0.0729571	0.07535941	0.15088356			6	3	0.29920006	0.09973335	0.0019637			
603-7	0.10544932	0.07521551	0.10406651			7	3	0.28473133	0.09491044	0.0002914			
603-8	0.13631529	0.07881258	0.10191226			8	3	0.31704013	0.10568004	0.00083729			
603-9	0.10380829	0.14514837	0.1185599			9	3	0.36751656	0.12250552	0.00043893			
603-10	0.07535322	0.09175984	0.08095523			10	3	0.24806828	0.08268943	6.955E-05			
603-11	0.0946874	0.09425163	0.09666569			11	3	0.28560472	0.09520157	1.6552E-06			
603-12	0.07245515	0.09750665	0.07939621			12	3	0.24935801	0.08311934	0.00016729			
603-13	0.13035754	0.12196369	0.10458498			13	3	0.3569062	0.11896873	0.00017278			
603-14	0.11213144	0.09257106	0.1105968			14	3	0.3152993	0.10509977	0.00011832			
603-15	0.09134776	0.08682176	0.11916561			15	3	0.29733513	0.09911171	0.00030674			
									0.1030805				
						ANOVA							
						<i>Source of variation</i>	<i>sums of squares (SS)</i>	<i>degrees of freedom (df)</i>	<i>Mean squares (MS)</i>	<i>F-value</i>	<i>P-value</i>	<i>critical F-value</i>	
						Between groups	0.00785018	14	0.00056073	0.76074717	0.69987411	2.03742044	
						Within groups	0.02211223	30	0.00073707				
						Total	0.02996241	44					
						within-sd	0.027149			status:	homogeneous		
						effective n	4.00						
						s_bb	0						
						s_bb_min	0.006898						
						u_bb	0.006898						
						Contribution to combined uncertainty:	0						

Cobalt:

	1	2	3									
603-1	0.04168381	0.04436165	0.03820285			<i>Bottle</i>	<i>Number</i>	<i>Sum</i>	<i>Mean</i>	<i>Variance</i>		
603-2	0.04224407	0.03794271	0.03953527			1	3	0.12424831	0.0414161	9.5364E-06		
603-3	0.04274222	0.03974967	0.04062967			2	3	0.11972205	0.03990735	4.7293E-06		
603-4	0.04222882	0.03857188	0.04196084			3	2	0.12312155	0.04104052	2.3654E-06		
603-5	0.0413699	0.04043464	0.03981146			4	3	0.12276154	0.04092051	4.155E-06		
603-6	0.03842189	0.04114737	0.03953129			5	2	0.121616	0.04053867	6.153E-07		
603-7	0.03959513	0.04009234	0.04087184			6	2	0.11910055	0.03970018	1.8784E-06		
603-8	0.03810101	0.04111654	0.03978305			7	3	0.12055931	0.04018644	4.1414E-07		
603-9	0.04040482	0.03805293	0.03775766			8	2	0.11900059	0.03966686	2.2835E-06		
603-10	0.03993063	0.04087401	0.04087938			9	3	0.11621541	0.03873847	2.1043E-06		
603-11	0.03993396	0.03924837	0.03856181			10	3	0.12168403	0.04056134	2.9835E-07		
603-12	0.04325771	0.03841336	0.0413948			11	3	0.11774413	0.03924804	4.707E-07		
603-13	0.03979537	0.04013191	0.04033948			12	3	0.12306587	0.04102196	5.9712E-06		
603-14	0.03949196	0.04033071	0.04160485			13	3	0.12026677	0.04008892	7.5401E-08		
603-15	0.04050478	0.04468323	0.04194216			14	2	0.12142751	0.04047584	1.1319E-06		
						15	3	0.12713017	0.04237672	4.5065E-06		
									0.04039253			
						ANOVA						
						<i>Source of variation</i>	<i>sums of squares (SS)</i>	<i>degrees of freedom (df)</i>	<i>Mean squares (MS)</i>	<i>F-value</i>	<i>P-value</i>	<i>critical F-value</i>
						Between groups	3.4674E-05	14	2.4767E-06	0.91649869	0.55223879	2.03742044
						Within groups	8.1072E-05	30	2.7024E-06			
						Total	0.00011575	44				
						within-sd	0.001644			status:	homogeneous	
						effective n	4.00					
						s_bb	0					
						s_bb_min	0.000418					
						u_bb	0.000418					
						Contribution to combined uncertainty:	0					